

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/28/2011

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.39: 0.078
LOT #: 02511080A2
EXPIRATION: 02/01/2013

BLANK TEST	0.000	11:42
INTERNAL STANDARD	VERIFIED	11:42
EXTERNAL STANDARD	0.078	11:42
BLANK TEST	0.000	11:43
EXTERNAL STANDARD	0.078	11:43
BLANK TEST	0.000	11:44
EXTERNAL STANDARD	0.078	11:44
BLANK TEST	0.000	11:45
EXTERNAL STANDARD	0.078	11:45
BLANK TEST	0.000	11:46
EXTERNAL STANDARD	0.078	11:46
BLANK TEST	0.000	11:47

Average = 0.0780
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 1.01
PIC: 3.02
Modem: 2.0
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 45.4°C	PASSED

PUMP INFO

Flow Rate = 5.441 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED


FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 20 11


Nikki Roth
Notary Public's Signature
My Commission Expires on with office

Nikki Roth
Notary Name




7/6/11